



Analyte solution
Congener: 2,3,7,8 TCDD

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	1,9	-0,16				
4	2,0	0,17				
6	2,0	0,030				
11	1,9	-0,27				
14	2,1	0,38				
18	2,3	0,84				
20	1,9	-0,071				
22	2,1	0,40				
28	2,0	-0,021				
30	1,8	-0,45				
32	1,6	-0,82				
35	1,8	-0,31				
39	2,1	0,39				
40	2,1	0,29				
43	2,3	0,97				
45	2,0	0,021				
48	19	44	Outlier			
54	1,9	-0,21				
58	1,9	-0,21				
62	2,0	0,091				
65	2,0	0,17				
76	2,4	1,1				
79	1,5	-1,2				
80	1,9	-0,055				
84	2,0	0,12				
86	1,9	-0,26				
90	1,9	-0,13				
93	1,8	-0,49				
98	1,9	-0,25				
100	2,1	0,28				
106	2,6	1,5				



Consensus statistics	
Consensus median, pg/g	2,0
Median all values pg/g	2,0
Consensus mean, pg/g	2,0
Standard deviation, pg/g	0,21
Relative standard deviation, %	11
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,7,8 PeCDD

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,1	0,065				
4	5,0	-0,0050				
6	5,3	0,31				
11	4,9	-0,094				
14	5,0	0,0050				
18	5,2	0,17				
20	5,2	0,20				
22	5,0	0,0050				
28	4,5	-0,56				
30	5,1	0,033				
32	5,4	0,33				
35	4,9	-0,16				
39	5,5	0,51				
40	4,8	-0,23				
43	5,4	0,42				
45	5,0	-0,0050				
48	54	49	Outlier			
54	5,0	0,015				
58	5,0	-0,045				
62	5,0	-0,044				
65	5,0	-0,038				
76	5,8	0,78				
79	4,7	-0,35				
80	5,0	-0,055				
84	5,0	-0,044				
86	5,1	0,065				
90	5,0	-0,0050				
93	5,1	0,025				
98	4,9	-0,17				
100	4,9	-0,16				
106	5,6	0,55				



Consensus statistics	
Consensus median, pg/g	5,0
Median all values pg/g	5,0
Consensus mean, pg/g	5,1
Standard deviation, pg/g	0,27
Relative standard deviation, %	5
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,4,7,8 HxCDD

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,1	0,095				
4	5,0	-0,025				
6	5,1	0,14				
11	4,5	-0,46				
14	5,3	0,27				
18	5,5	0,53				
20	5,2	0,23				
22	5,1	0,12				
28	4,9	-0,12				
30	4,5	-0,52				
32	4,5	-0,48				
35	5,3	0,28				
39	5,0	-0,018				
40	4,7	-0,34				
43	5,8	0,81				
45	4,6	-0,41				
48	58	53	Outlier			
54	5,0	-0,055				
58	4,9	-0,075				
62	5,2	0,18				
65	5,0	-0,0017				
76	5,9	0,86				
79	4,3	-0,74				
80	5,1	0,075				
84	5,0	0,0017				
86	4,8	-0,21				
90	4,4	-0,64				
93	4,9	-0,075				
98	5,0	0,0067				
100	5,2	0,16				
106	5,7	0,70				



Consensus statistics	
Consensus median, pg/g	5,0
Median all values pg/g	5,0
Consensus mean, pg/g	5,0
Standard deviation, pg/g	0,40
Relative standard deviation, %	8
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,6,7,8 HxCDD

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	4,9	0,14				
4	5,0	0,27				
6	5,3	0,58				
11	4,5	-0,34				
14	4,8	0,015				
18	5,3	0,54				
20	5,0	0,21				
22	4,8	-0,016				
28	4,7	-0,063				
30	4,4	-0,41				
32	4,3	-0,52				
35	4,4	-0,39				
39	4,5	-0,29				
40	4,7	-0,11				
43	5,1	0,36				
45	4,6	-0,25				
48	44	41	Outlier			
54	4,8	-0,027				
58	4,8	0,036				
62	4,8	-0,023				
65	4,9	0,14				
76	5,6	0,81				
79	4,0	-0,82				
80	4,9	0,14				
84	5,0	0,20				
86	4,9	0,14				
90	4,3	-0,53				
93	4,6	-0,16				
98	4,8	-0,015				
100	5,0	0,18				
106	5,4	0,68				



Consensus statistics	
Consensus median, pg/g	4,8
Median all values pg/g	4,8
Consensus mean, pg/g	4,8
Standard deviation, pg/g	0,35
Relative standard deviation, %	7
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,7,8,9 HxCDD

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,1	0,11	Outlier			
4	5,0	0,066				
6	5,3	0,41				
11	4,5	-0,43				
14	4,8	-0,15				
18	5,5	0,52				
20	5,1	0,16				
22	5,0	0,015				
28	4,5	-0,41				
30	4,8	-0,18				
32	4,9	-0,064				
35	4,2	-0,78				
39	5,4	0,49				
40	4,3	-0,61				
43	5,2	0,24				
45	4,8	-0,16				
48	5,7	0,52				
54	5,0	0,015				
58	4,9	-0,015				
62	4,8	-0,15				
65	5,0	0,047				
76	5,7	0,74				
79	4,0	-0,98				
80	4,8	-0,11				
84	5,0	0,064				
86	5,2	0,25				
90	4,8	-0,17				
93	5,1	0,11				
98	4,7	-0,23				
100	4,9	-0,035				
106	5,5	0,53				



Consensus statistics	
Consensus median, pg/g	4,9
Median all values pg/g	5,0
Consensus mean, pg/g	4,9
Standard deviation, pg/g	0,38
Relative standard deviation, %	8
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,4,6,7,8 HpCDD

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,1	0,077				
4	5,0	0,027				
6	5,1	0,16				
11	4,8	-0,15				
14	4,9	-0,14				
18	5,9	0,94				
20	5,1	0,10				
22	4,8	-0,21				
28	4,8	-0,15				
30	4,3	-0,67				
32	4,6	-0,39				
35	4,8	-0,17				
39	4,8	-0,18				
40	4,5	-0,53				
43	5,0	0,015				
45	4,6	-0,38				
48	53	48	Outlier			
54	5,2	0,19				
58	5,2	0,17				
62	4,7	-0,24				
65	5,0	0,0021				
76	5,9	0,90				
79	4,8	-0,20				
80	4,8	-0,16				
84	5,0	-0,0021				
86	4,8	-0,19				
90	5,0	0,027				
93	5,2	0,22				
98	5,1	0,16				
100	5,1	0,14				
106	5,1	0,057				

Consensus statistics	
Consensus median, pg/g	5,0
Median all values pg/g	5,0
Consensus mean, pg/g	5,0
Standard deviation, pg/g	0,33
Relative standard deviation, %	7
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0





Analyte solution

Congener: OCDD

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	10	0,21				
4	9,9	0,15				
6	10	0,25				
11	10	0,23				
14	8,9	-0,38				
18	11	0,64				
20	9,4	-0,11				
22	9,5	-0,079				
28	9,6	-0,0035				
30	8,9	-0,39				
32	9,1	-0,28				
35	9,6	-0,053				
39	9,2	-0,24				
40	9,8	0,055				
43	9,8	0,087				
45	9,2	-0,25				
48	111	52	Outlier			
54	9,6	-0,022				
58	9,7	0,0035				
62	9,4	-0,13				
65	9,9	0,15				
76	10	0,28				
79	7,5	-1,1				
80	9,5	-0,064				
84	10	0,18				
86	9,7	0,011				
90	10	0,28				
93	9,5	-0,059				
98	9,0	-0,34				
100	10	0,22				
106	10	0,28				

Consensus statistics

Consensus median, pg/g	9,7
Median all values pg/g	9,7
Consensus mean, pg/g	9,6
Standard deviation, pg/g	0,61
Relative standard deviation, %	6
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0





Analyte solution

Congener: 2,3,7,8 TCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	2,1	0,15				
4	2,0	0,075				
6	2,0	0,060				
11	1,9	-0,23				
14	1,9	-0,20				
18	2,5	1,2				
20	2,0	0,0000				
22	2,0	0,10				
28	1,9	-0,12				
30	1,8	-0,50				
32	1,7	-0,71				
35	2,1	0,26				
39	2,1	0,22				
40	1,8	-0,37				
43	2,2	0,45				
45	2,0	0,0000				
48	22	49	Outlier			
54	2,0	-0,050				
58	2,0	-0,075				
62	2,1	0,24				
65	2,0	0,15				
76	2,1	0,38				
79	1,6	-1,1				
80	2,1	0,25				
84	2,0	-0,0038				
86	1,8	-0,37				
90	1,7	-0,65				
93	2,0	-0,10				
98	2,2	0,44				
100	2,0	0,025				
106	1,9	-0,13				

Consensus statistics



Consensus median, pg/g	2,0
Median all values pg/g	2,0
Consensus mean, pg/g	2,0
Standard deviation, pg/g	0,16
Relative standard deviation, %	8
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,7,8 PeCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,0	0,015				
4	5,2	0,23				
6	5,0	0,077				
11	4,7	-0,25				
14	4,7	-0,26				
18	5,8	0,84				
20	5,1	0,11				
22	4,8	-0,12				
28	4,6	-0,39				
30	4,8	-0,18				
32	4,6	-0,38				
35	5,1	0,17				
39	5,2	0,21				
40	4,6	-0,31				
43	5,2	0,24				
45	4,6	-0,34				
48	54	49	Outlier			
54	5,1	0,12				
58	4,9	-0,015				
62	4,8	-0,15				
65	4,8	-0,13				
76	5,7	0,79				
79	4,6	-0,41				
80	4,8	-0,19				
84	5,0	0,037				
86	5,0	0,062				
90	4,8	-0,13				
93	5,1	0,17				
98	5,4	0,49				
100	5,2	0,22				
106	4,8	-0,21				



Consensus statistics	
Consensus median, pg/g	5,0
Median all values pg/g	5,0
Consensus mean, pg/g	5,0
Standard deviation, pg/g	0,31
Relative standard deviation, %	6
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 2,3,4,7,8 PeCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,2	0,19				
4	5,0	0,0064				
6	5,3	0,27				
11	4,5	-0,54				
14	4,7	-0,29				
18	5,1	0,13				
20	4,7	-0,31				
22	5,0	0,016				
28	5,1	0,067				
30	4,5	-0,47				
32	4,5	-0,51				
35	5,0	0,037				
39	5,0	-0,041				
40	4,7	-0,28				
43	5,4	0,44				
45	4,6	-0,44				
48	49	44	Outlier			
54	4,7	-0,27				
58	4,7	-0,26				
62	5,1	0,063				
65	5,0	0,0015				
76	5,3	0,34				
79	4,7	-0,30				
80	5,1	0,14				
84	5,0	-0,0015				
86	5,0	0,019				
90	4,2	-0,76				
93	4,7	-0,31				
98	5,2	0,22				
100	5,1	0,10				
106	4,9	-0,15				

Consensus statistics	
Consensus median, pg/g	5,0
Median all values pg/g	5,0
Consensus mean, pg/g	4,9
Standard deviation, pg/g	0,29
Relative standard deviation, %	6
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0





Analyte solution
Congener: 1,2,3,4,7,8 HxCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,1	0,11				
4	5,2	0,24				
6	5,3	0,28				
11	4,7	-0,28				
14	5,1	0,14				
18	5,3	0,35				
20	5,0	-0,025				
22	5,0	-0,025				
28	4,7	-0,25				
30	4,6	-0,43				
32	4,6	-0,41				
35	4,5	-0,49				
39	4,6	-0,38				
40	4,6	-0,37				
43	5,4	0,39				
45	4,8	-0,18				
48	50	45	Outlier			
54	5,1	0,15				
58	5,1	0,14				
62	4,9	-0,068				
65	5,1	0,10				
76	5,2	0,23				
79	4,1	-0,89				
80	4,9	-0,075				
84	5,0	0,0048				
86	4,6	-0,35				
90	5,0	-0,0048				
93	5,3	0,29				
98	5,3	0,33				
100	5,1	0,12				
106	5,3	0,36				

Consensus statistics



Consensus median, pg/g	5,0
Median all values pg/g	5,0
Consensus mean, pg/g	4,9
Standard deviation, pg/g	0,31
Relative standard deviation, %	6
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,6,7,8 HxCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,1	0,083				
4	5,1	0,063				
6	5,3	0,29				
11	4,8	-0,19				
14	5,1	0,073				
18	5,9	0,91				
20	5,1	0,083				
22	5,1	0,083				
28	4,5	-0,48				
30	4,7	-0,30				
32	4,6	-0,36				
35	5,0	-0,0029				
39	5,3	0,32				
40	4,7	-0,26				
43	5,2	0,17				
45	4,8	-0,16				
48	4,6	41	Outlier			
54	4,8	-0,18				
58	4,9	-0,11				
62	5,1	0,11				
65	5,0	0,045				
76	5,6	0,65				
79	4,5	-0,49				
80	5,2	0,18				
84	5,0	0,0029				
86	5,0	-0,016				
90	4,9	-0,14				
93	4,9	-0,057				
98	4,9	-0,063				
100	4,9	-0,13				
106	5,2	0,19				



Consensus statistics	
Consensus median, pg/g	5,0
Median all values pg/g	5,0
Consensus mean, pg/g	5,0
Standard deviation, pg/g	0,29
Relative standard deviation, %	6
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 2,3,4,6,7,8 HxCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,2	0,25				
4	5,0	0,081				
6	5,1	0,23				
11	4,6	-0,29				
14	4,7	-0,20				
18	5,2	0,29				
20	5,0	0,061				
22	4,9	0,010				
28	4,9	-0,029				
30	4,4	-0,53				
32	4,6	-0,34				
35	4,9	0,016				
39	5,4	0,48				
40	4,5	-0,44				
43	4,9	-0,019				
45	4,7	-0,17				
48	44	40	Outlier			
54	4,9	0,020				
58	4,9	-0,010				
62	4,9	-0,015				
65	5,1	0,18				
76	5,3	0,35				
79	4,5	-0,41				
80	4,9	-0,031				
84	5,0	0,088				
86	4,9	-0,045				
90	4,8	-0,092				
93	4,8	-0,081				
98	5,1	0,22				
100	5,0	0,13				
106	5,4	0,49				



Consensus statistics	
Consensus median, pg/g	4,9
Median all values pg/g	4,9
Consensus mean, pg/g	4,9
Standard deviation, pg/g	0,25
Relative standard deviation, %	5
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,7,8,9 HxCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,2	0,14				
4	5,1	0,080				
6	5,2	0,18				
11	5,0	-0,026				
14	4,8	-0,24				
18	5,7	0,66				
20	5,1	0,10				
22	4,9	-0,15				
28	4,9	-0,11				
30	4,5	-0,53				
32	5,0	0,030				
35	4,8	-0,20				
39	4,8	-0,22				
40	4,7	-0,29				
43	5,0	-0,049				
45	4,7	-0,27				
48	5,2	0,47	Outlier			
54	5,1	0,060				
58	5,1	0,060				
62	4,9	-0,15				
65	5,0	0,034				
76	5,8	0,76				
79	4,7	-0,34				
80	4,9	-0,089				
84	5,0	-0,020				
86	5,0	-0,048				
90	5,1	0,090				
93	5,3	0,29				
98	5,2	0,21				
100	5,0	0,020				
106	5,3	0,30				



Consensus statistics	
Consensus median, pg/g	5,0
Median all values pg/g	5,0
Consensus mean, pg/g	5,0
Standard deviation, pg/g	0,27
Relative standard deviation, %	5
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,4,6,7,8 HpCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,2	0,18				
4	5,3	0,32				
6	5,2	0,15				
11	5,1	0,060				
14	5,0	0,0000				
18	5,6	0,60				
20	5,1	0,070				
22	5,1	0,10				
28	5,0	-0,016				
30	4,6	-0,44				
32	5,2	0,20				
35	5,0	-0,022				
39	4,5	-0,46				
40	4,8	-0,21				
43	5,4	0,39				
45	4,5	-0,46				
48	50	45	Outlier			
54	4,8	-0,20				
58	4,8	-0,22				
62	5,1	0,12				
65	4,9	-0,087				
76	5,7	0,72				
79	4,0	-0,98				
80	5,0	-0,040				
84	5,0	-0,0040				
86	4,8	-0,21				
90	4,7	-0,29				
93	5,3	0,32				
98	5,9	0,86				
100	5,0	0,030				
106	5,0	0,0000				



Consensus statistics	
Consensus median, pg/g	5,0
Median all values pg/g	5,0
Consensus mean, pg/g	5,0
Standard deviation, pg/g	0,37
Relative standard deviation, %	7
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0



Analyte solution
Congener: 1,2,3,4,7,8,9 HpCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	5,1	0,19				
4	5,2	0,34				
6	5,2	0,34				
11	5,3	0,40				
14	4,6	-0,35				
18	5,5	0,57				
20	4,9	-0,010				
22	4,9	0,010				
28	4,9	-0,047				
30	4,4	-0,54				
32	5,4	0,47				
35	4,6	-0,32				
39	4,6	-0,27				
40	4,5	-0,41				
43	5,2	0,32				
45	4,6	-0,33				
48	49	45	Outlier			
54	4,8	-0,071				
58	4,8	-0,10				
62	4,8	-0,064				
65	5,0	0,13				
76	5,9	1,0				
79	4,1	-0,81				
80	5,0	0,051				
84	5,0	0,092				
86	4,8	-0,14				
90	4,4	-0,53				
93	5,6	0,66				
98	5,3	0,44				
100	5,0	0,051				
106	4,9	-0,010				

Consensus statistics	
Consensus median, pg/g	4,9
Median all values pg/g	4,9
Consensus mean, pg/g	4,9
Standard deviation, pg/g	0,39
Relative standard deviation, %	8
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0





Analyte solution

Congener: OCDF

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	10	0,050				
4	10	0,050				
6	10	0,13				
11	10	0,018				
14	10	0,0050				
18	8,9	-0,56				
20	10	0,12				
22	10	-0,0050				
28	9,5	-0,22				
30	9,0	-0,52				
32	8,8	-0,58				
35	9,3	-0,37				
39	9,8	-0,11				
40	9,4	-0,30				
43	11	0,33				
45	9,1	-0,46				
48	107	49	Outlier			
54	9,8	-0,080				
58	9,8	-0,090				
62	10	0,0065				
65	10	0,033				
76	10	0,25				
79	8,1	-0,97				
80	10	0,075				
84	10	0,0080				
86	9,8	-0,090				
90	11	0,41				
93	9,9	-0,055				
98	12	0,87				
100	9,9	-0,040				
106	10	0,0050				

Consensus statistics

Consensus median, pg/g	10
Median all values pg/g	10
Consensus mean, pg/g	9,9
Standard deviation, pg/g	0,68
Relative standard deviation, %	7
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0





Analyte solution

Congener: PCB 77

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	8,5	-0,78				
4	10	0,11				
6	10	-0,062				
11	10	0,064				
14	9,1	-0,48				
18	10	-0,014				
20	10	-0,019				
22	10	-0,0037				
28	9,7	-0,21				
30	9,6	-0,26				
32	9,4	-0,34				
35	9,3	-0,38				
39	10	0,048				
40	10	0,13				
43	10	0,016				
45	11	0,20				
48	93	41	Outlier			
54	11	0,20				
58	10	0,11				
62	10	0,018				
65	10	-0,0060				
76	9,6	-0,24				
79	11	0,55				
80	10	0,15				
84	10	0,0037				
86	11	0,25				
90	10	0,056				
93	10	-0,053				
98	9,7	-0,19				
100	10	0,11				
106	10	-0,043				

Consensus statistics

Consensus median, pg/g	10
Median all values pg/g	10
Consensus mean, pg/g	10
Standard deviation, pg/g	0,50
Relative standard deviation, %	5
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0





Analyte solution

Congener: PCB 126

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	9,1	-0,42				
4	10	0,038				
6	10	0,22				
11	10	0,14				
14	10	0,038				
18	10	0,27				
20	9,1	-0,40				
22	11	0,60				
28	9,1	-0,41				
30	9,3	-0,33				
32	9,5	-0,24				
35	9,2	-0,37				
39	9,8	-0,082				
40	10	0,084				
43	12	0,93				
45	10	0,17				
48	91	41	Outlier			
54	10,0	0,028				
58	9,8	-0,063				
62	11	0,59				
65	10	0,11				
76	9,6	-0,15				
79	9,8	-0,074				
80	11	0,51				
84	10	0,065				
86	9,8	-0,059				
90	9,2	-0,35				
93	9,2	-0,36				
98	9,5	-0,22				
100	11	0,69				
106	9,9	-0,028				

Consensus statistics

Consensus median, pg/g	9,9
Median all values pg/g	10
Consensus mean, pg/g	10
Standard deviation, pg/g	0,70
Relative standard deviation, %	7
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0





Analyte solution

Congener: PCB 169

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	8,6	-0,65				
4	10	0,12				
6	10	0,14				
11	11	0,33				
14	9,3	-0,29				
18	9,4	-0,26				
20	9,6	-0,13				
22	9,9	0,0095				
28	9,9	-0,011				
30	10	0,20				
32	8,9	-0,50				
35	9,0	-0,42				
39	9,2	-0,35				
40	10	0,13				
43	10	0,22				
45	11	0,72				
48	95	43	Outlier			
54	10	0,12				
58	9,8	-0,031				
62	10	0,057				
65	10	0,063				
76	9,7	-0,11				
79	9,8	-0,044				
80	11	0,54				
84	10	0,078				
86	9,9	-0,0095				
90	9,7	-0,092				
93	9,5	-0,18				
98	8,9	-0,49				
100	10	0,12				
106	10	0,12				

Consensus statistics

Consensus median, pg/g	9,9
Median all values pg/g	9,9
Consensus mean, pg/g	9,8
Standard deviation, pg/g	0,58
Relative standard deviation, %	6
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0





Analyte solution

Congener: PCB 81

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	8,5	-0,74				
4	10	0,20				
6	10	0,016				
11	10	-0,014				
14	9,8	-0,13				
18	11	0,30				
20	9,5	-0,25				
22	10	0,010				
28	9,5	-0,25				
30	9,6	-0,20				
32	9,2	-0,38				
35	9,2	-0,39				
39	9,5	-0,27				
40	10	0,23				
43	11	0,41				
45	11	0,46				
48	96	43	Outlier			
54	9,8	-0,13				
58	10	0,10				
62	10	0,047				
65	10	0,056				
76	10	0,064				
79	11	0,30				
80	10	0,12				
84	10	-0,0065				
86	10	0,000025				
90	9,1	-0,44				
93	11	0,36				
98	9,8	-0,11				
100	9,2	-0,40				
106	10	-0,000025				

Consensus statistics

Consensus median, pg/g	10
Median all values pg/g	10
Consensus mean, pg/g	9,9
Standard deviation, pg/g	0,56
Relative standard deviation, %	6
No. of values reported	31
No. of values removed	1
No. of reported non-detects	0





Analyte solution

Congener: PCB 105

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	94	-0,25				
4	101	0,12				
6	100	0,060				
11	100	0,090				
14	99	0,0052				
18	96	-0,12				
20	100	0,071				
22	95	-0,17				
28	91	-0,37				
30	94	-0,21				
32	99	-0,0052				
35	112	0,66				
39	94	-0,23				
40	95	-0,17				
43	96	-0,14				
45	105	0,32				
54	96	-0,16				
58	96	-0,14				
62	96	-0,15				
65	102	0,17				
76	96	-0,12				
79	109	0,53				
80	97	-0,10				
84	100	0,051				
86	101	0,094				
90	95	-0,18				
93	104	0,27				
98	100	0,090				
100	102	0,17				
106	100	0,067				

Consensus statistics

Consensus median, pg/g	99
Median all values pg/g	99
Consensus mean, pg/g	99
Standard deviation, pg/g	4,5
Relative standard deviation, %	5
No. of values reported	30
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PCB 114

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	89	-0,52				
4	99	-0,019				
6	100	0,027				
11	100	0,020				
14	103	0,14				
18	95	-0,22				
20	106	0,30				
22	105	0,27				
28	94	-0,29				
30	94	-0,30				
32	100	-0,0014				
35	110	0,50				
39	99	-0,053				
40	97	-0,13				
43	104	0,19				
45	96	-0,17				
54	97	-0,14				
58	96	-0,17				
62	105	0,28				
65	99	-0,025				
76	100	0,0014				
79	111	0,59				
80	106	0,30				
84	100	0,0080				
86	100	-0,0070				
90	97	-0,16				
93	101	0,049				
98	101	0,066				
100	101	0,064				
106	100	-0,012				

Consensus statistics

Consensus median, pg/g	100
Median all values pg/g	100
Consensus mean, pg/g	100
Standard deviation, pg/g	4,7
Relative standard deviation, %	5
No. of values reported	30
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PCB 118

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	89	-0,56				
4	99	-0,092				
6	101	0,039				
11	102	0,089				
14	98	-0,11				
18	93	-0,39				
20	102	0,071				
22	96	-0,22				
28	102	0,047				
30	89	-0,56				
32	103	0,10				
35	99	-0,087				
39	96	-0,24				
40	101	0,025				
43	106	0,25				
45	111	0,52				
54	101	0,017				
58	100	-0,032				
62	96	-0,22				
65	102	0,089				
76	98	-0,14				
79	112	0,54				
80	95	-0,26				
84	100	-0,034				
86	99	-0,059				
90	106	0,27				
93	106	0,25				
98	100	-0,017				
100	110	0,46				
106	101	0,017				

Consensus statistics

Consensus median, pg/g	101
Median all values pg/g	101
Consensus mean, pg/g	100
Standard deviation, pg/g	5,4
Relative standard deviation, %	5
No. of values reported	30
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PCB 123

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	98	-0,13				
4	101	0,0061				
6	101	-0,013				
11	103	0,12				
14	102	0,056				
18	97	-0,19				
20	102	0,077				
22	104	0,17				
28	99	-0,10				
30	92	-0,43				
32	101	0,030				
35	98	-0,14				
39	99	-0,088				
40	98	-0,15				
43	101	0,0032				
45	106	0,23				
54	100	-0,039				
58	97	-0,18				
62	104	0,18				
65	101	-0,0032				
76	97	-0,20				
79	111	0,52				
80	101	0,030				
84	100	-0,031				
86	101	0,025				
90	105	0,21				
93	100	-0,063				
98	99	-0,10				
100	106	0,26				
106	101	0,011				

Consensus statistics

Consensus median, pg/g	101
Median all values pg/g	101
Consensus mean, pg/g	101
Standard deviation, pg/g	3,6
Relative standard deviation, %	4
No. of values reported	30
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PCB 156

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	91	-0,41				
4	97	-0,11				
6	101	0,12				
11	101	0,090				
14	96	-0,16				
18	96	-0,16				
20	99	0,025				
22	100	0,071				
28	98	-0,025				
30	95	-0,19				
32	95	-0,20				
35	95	-0,17				
39	97	-0,10				
40	97	-0,11				
43	106	0,37				
45	100	0,081				
54	97	-0,10				
58	100	0,040				
62	100	0,078				
65	100	0,071				
76	101	0,086				
79	111	0,60				
80	98	-0,034				
84	100	0,061				
86	92	-0,32				
90	97	-0,10				
93	96	-0,16				
98	103	0,24				
100	105	0,31				
106	101	0,11				

Consensus statistics

Consensus median, pg/g	99
Median all values pg/g	99
Consensus mean, pg/g	99
Standard deviation, pg/g	4,1
Relative standard deviation, %	4
No. of values reported	30
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PCB 157

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	88	-0,54				
4	101	0,11				
6	100	0,076				
11	100	0,10				
14	96	-0,10				
18	94	-0,24				
20	94	-0,24				
22	98	0,0036				
28	101	0,13				
30	99	0,019				
32	93	-0,26				
35	97	-0,072				
39	97	-0,090				
40	96	-0,13				
43	106	0,39				
45	100	0,084				
54	96	-0,12				
58	96	-0,10				
62	98	-0,037				
65	99	0,018				
76	98	-0,0015				
79	111	0,64				
80	96	-0,13				
84	100	0,078				
86	98	0,0015				
90	91	-0,36				
93	106	0,38				
98	104	0,26				
100	107	0,44				
106	89	-0,45				

Consensus statistics

Consensus median, pg/g	98
Median all values pg/g	98
Consensus mean, pg/g	98
Standard deviation, pg/g	5,1
Relative standard deviation, %	5
No. of values reported	30
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PCB 167

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	89	-0,47				
4	99	0,040				
6	101	0,17				
11	98	0,013				
14	95	-0,15				
18	93	-0,26				
20	101	0,15				
22	97	-0,046				
28	98	-0,010				
30	87	-0,57				
32	95	-0,14				
35	97	-0,068				
39	95	-0,15				
40	96	-0,13				
43	101	0,16				
45	95	-0,17				
54	103	0,25				
58	98	0,0044				
62	97	-0,026				
65	98	-0,0044				
76	100	0,12				
79	114	0,81				
80	95	-0,14				
84	100	0,12				
86	101	0,15				
90	94	-0,21				
93	114	0,82				
98	108	0,52				
100	116	0,92				
106	100	0,10				

Consensus statistics

Consensus median, pg/g	98
Median all values pg/g	98
Consensus mean, pg/g	99
Standard deviation, pg/g	6,6
Relative standard deviation, %	7
No. of values reported	30
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PCB 189

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	91	-0,41				
4	96	-0,12				
6	101	0,11				
11	99	0,021				
14	96	-0,15				
18	75	-1,2				
20	98	-0,022				
22	100	0,080				
28	91	-0,37				
30	87	-0,60				
32	93	-0,29				
35	100	0,072				
39	99	-0,0010				
40	97	-0,080				
43	100	0,089				
45	99	0,0010				
54	102	0,17				
58	98	-0,029				
62	100	0,090				
65	100	0,083				
76	100	0,084				
79	111	0,63				
80	95	-0,18				
84	100	0,086				
86	100	0,067				
90	90	-0,44				
93	105	0,34				
98	98	-0,041				
100	110	0,57				
106	97	-0,070				

Consensus statistics

Consensus median, pg/g	99
Median all values pg/g	99
Consensus mean, pg/g	98
Standard deviation, pg/g	6,7
Relative standard deviation, %	7
No. of values reported	30
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PCB 28

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	100	-0,028				
4	100	-0,041				
6	101	0,013				
11	100	-0,013				
14	103	0,12				
18	81	-0,96				
22	101	0,016				
28	97	-0,18				
30	95	-0,27				
32	108	0,39				
35	91	-0,49				
39	100	-0,022				
40	107	0,31				
43	97	-0,19				
45	99	-0,076				
54	105	0,22				
58	105	0,22				
62	101	0,039				
65	103	0,13				
79	122	1,1				
80	103	0,13				
84	100	-0,025				
86	99	-0,093				
90	96	-0,21				
93	116	0,76				
98	108	0,39				
100	112	0,57				
106	99	-0,076				

Consensus statistics

Consensus median, pg/g	101
Median all values pg/g	101
Consensus mean, pg/g	102
Standard deviation, pg/g	7,6
Relative standard deviation, %	7
No. of values reported	28
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PCB 52

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	96	-0,22				
4	100	-0,037				
6	101	0,018				
11	104	0,18				
14	103	0,12				
18	74	-1,3				
22	99	-0,056				
28	99	-0,081				
30	90	-0,52				
32	106	0,30				
35	99	-0,078				
39	101	0,0080				
40	109	0,41				
43	98	-0,13				
45	101	0,033				
54	103	0,13				
58	104	0,18				
62	99	-0,064				
65	100	-0,0080				
79	106	0,30				
80	105	0,24				
84	100	-0,015				
86	101	0,047				
90	90	-0,53				
93	94	-0,30				
98	100	-0,039				
100	122	1,1				
106	103	0,13				

Consensus statistics



Consensus median, pg/g	100
Median all values pg/g	100
Consensus mean, pg/g	100
Standard deviation, pg/g	7,8
Relative standard deviation, %	8
No. of values reported	28
No. of values removed	0
No. of reported non-detects	0



Analyte solution
Congener: PCB 101

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	99	-0,072				
4	101	-0,0057				
6	101	0,011				
11	105	0,23				
14	103	0,12				
18	81	-0,97				
22	101	0,024				
28	97	-0,19				
30	92	-0,44				
32	97	-0,16				
35	96	-0,24				
39	98	-0,15				
40	101	0,0057				
43	112	0,54				
45	100	-0,036				
54	106	0,26				
58	102	0,064				
62	102	0,059				
65	102	0,078				
79	103	0,11				
80	105	0,23				
84	100	-0,033				
86	95	-0,28				
90	87	-0,69				
93	94	-0,33				
98	97	-0,18				
100	127	1,3				
106	105	0,21				



Consensus statistics	
Consensus median, pg/g	101
Median all values pg/g	101
Consensus mean, pg/g	100
Standard deviation, pg/g	8,0
Relative standard deviation, %	8
No. of values reported	28
No. of values removed	0
No. of reported non-detects	0



Analyte solution
Congener: PCB 138

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	97	-0,23				
4	97	-0,25				
6	100	-0,078				
11	104	0,094				
14	105	0,17				
18	75	-1,3				
22	104	0,12				
28	97	-0,26				
30	95	-0,34				
32	110	0,40				
35	94	-0,40				
39	99	-0,15				
40	107	0,24				
43	107	0,26				
45	102	0,0053				
54	103	0,054				
58	101	-0,044				
62	104	0,12				
65	105	0,16				
79	106	0,21				
80	110	0,41				
84	100	-0,071				
86	102	-0,0053				
90	94	-0,39				
93	101	-0,024				
98	97	-0,25				
100	129	1,3				
106	103	0,054				



Consensus statistics	
Consensus median, pg/g	102
Median all values pg/g	102
Consensus mean, pg/g	102
Standard deviation, pg/g	8,6
Relative standard deviation, %	8
No. of values reported	28
No. of values removed	0
No. of reported non-detects	0



Analyte solution
Congener: PCB 153

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	95	-0,29				
4	96	-0,27				
6	101	-0,025				
11	104	0,14				
14	102	0,025				
18	88	-0,67				
22	103	0,10				
28	98	-0,16				
30	94	-0,36				
32	109	0,38				
35	96	-0,26				
39	98	-0,17				
40	106	0,23				
43	97	-0,20				
45	102	0,035				
54	104	0,13				
58	105	0,18				
62	103	0,098				
65	105	0,18				
79	103	0,092				
80	107	0,27				
84	100	-0,049				
86	101	-0,030				
90	93	-0,42				
93	98	-0,17				
98	94	-0,34				
100	129	1,4				
106	103	0,084				



Consensus statistics	
Consensus median, pg/g	101
Median all values pg/g	101
Consensus mean, pg/g	101
Standard deviation, pg/g	7,3
Relative standard deviation, %	7
No. of values reported	28
No. of values removed	0
No. of reported non-detects	0



Analyte solution
Congener: PCB 180

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	95	-0,28				
4	98	-0,14				
6	101	-0,016				
11	101	-0,0074				
14	104	0,18				
18	78	-1,1				
22	104	0,13				
28	95	-0,30				
30	93	-0,37				
32	102	0,043				
35	96	-0,23				
39	98	-0,14				
40	105	0,21				
43	111	0,50				
45	101	0,0074				
54	106	0,26				
58	106	0,26				
62	103	0,11				
65	104	0,17				
79	108	0,33				
80	107	0,29				
84	100	-0,049				
86	95	-0,27				
90	97	-0,21				
93	96	-0,26				
98	97	-0,19				
100	114	0,65				
106	106	0,26				


Consensus statistics	
Consensus median, pg/g	101
Median all values pg/g	101
Consensus mean, pg/g	101
Standard deviation, pg/g	6,9
Relative standard deviation, %	7
No. of values reported	28
No. of values removed	0
No. of reported non-detects	0



Analyte solution
Congener: PBDE 28

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	24	-0,18				
4	25	0,0000				
11	26	0,16				
20	24	-0,34				
22	25	-0,016				
28	27	0,45				
35	24	-0,20				
39	27	0,43				
43	25	-0,044				
62	26	0,10				
76	25	0,054				
79	24	-0,32				
80	28	0,55				
90	23	-0,42				
93	23	-0,42				
100	27	0,26				
106	26	0,079				

Consensus statistics	
Consensus median, pg/g	25
Median all values pg/g	25
Consensus mean, pg/g	25
Standard deviation, pg/g	1,5
Relative standard deviation, %	6
No. of values reported	17
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PBDE 47

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	24	-0,17				
4	25	0,13				
11	26	0,34				
20	23	-0,28				
22	24	0,0000				
28	27	0,54				
35	24	-0,16				
39	26	0,40				
43	24	0,013				
62	24	-0,17				
76	26	0,37				
79	23	-0,28				
80	24	-0,023				
90	23	-0,26				
93	23	-0,28				
100	26	0,27				
106	25	0,15				

Consensus statistics

Consensus median, pg/g	24
Median all values pg/g	24
Consensus mean, pg/g	25
Standard deviation, pg/g	1,3
Relative standard deviation, %	5
No. of values reported	17
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PBDE 99

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	25	-0,020				
4	25	0,079				
11	27	0,47				
20	25	0,13				
22	24	-0,14				
28	30	1,1				
35	24	-0,072				
39	28	0,66				
43	23	-0,36				
62	24	-0,16				
76	24	-0,066				
79	21	-0,68				
80	25	0,0000				
90	23	-0,42				
93	25	0,030				
100	26	0,17				
106	25	0,11				

Consensus statistics

Consensus median, pg/g	25
Median all values pg/g	25
Consensus mean, pg/g	25
Standard deviation, pg/g	2,1
Relative standard deviation, %	8
No. of values reported	17
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PBDE 100

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	22	-0,61				
4	25	0,089				
11	26	0,25				
20	25	0,0000				
22	24	-0,13				
28	31	1,3				
35	24	-0,13				
39	27	0,48				
43	25	0,10				
62	24	-0,13				
76	24	-0,083				
79	23	-0,25				
80	25	0,026				
90	21	-0,69				
93	23	-0,28				
100	27	0,38				
106	25	0,061				

Consensus statistics

Consensus median, pg/g	25
Median all values pg/g	25
Consensus mean, pg/g	25
Standard deviation, pg/g	2,2
Relative standard deviation, %	9
No. of values reported	17
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PBDE 153

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	23	-0,29				
4	25	0,11				
11	25	0,13				
20	21	-0,72				
22	24	-0,11				
28	30	1,0				
35	22	-0,58				
39	27	0,42				
43	25	0,15				
62	25	0,0000				
76	24	-0,11				
79	24	-0,13				
80	25	0,029				
90	27	0,47				
93	23	-0,34				
100	25	0,11				
106	25	-0,016				

Consensus statistics

Consensus median, pg/g	25
Median all values pg/g	25
Consensus mean, pg/g	25
Standard deviation, pg/g	2,0
Relative standard deviation, %	8
No. of values reported	17
No. of values removed	0
No. of reported non-detects	0




Analyte solution

Congener: PBDE 154

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	23	-0,20				
4	25	0,16				
11	25	0,19				
20	23	-0,30				
22	25	0,12				
28	32	1,6				
35	22	-0,50				
39	26	0,31				
43	26	0,37				
62	24	0,0000				
76	24	-0,0057				
79	22	-0,40				
80	25	0,16				
90	23	-0,24				
93	23	-0,36				
100	26	0,36				
106	24	-0,055				

Consensus statistics	
Consensus median, pg/g	24
Median all values pg/g	24
Consensus mean, pg/g	25
Standard deviation, pg/g	2,3
Relative standard deviation, %	9
No. of values reported	17
No. of values removed	0
No. of reported non-detects	0





Analyte solution

Congener: PBDE 183

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	19	-1,0				
4	25	0,23				
11	25	0,18				
20	24	0,037				
22	25	0,18				
28	34	2,1				
35	23	-0,23				
39	24	-0,036				
43	22	-0,52				
62	24	-0,013				
76	24	0,025				
79	22	-0,49				
80	24	0,0000				
90	20	-0,77				
93	22	-0,42				
100	26	0,29				
106	25	0,12				

Consensus statistics

Consensus median, pg/g	24
Median all values pg/g	24
Consensus mean, pg/g	24
Standard deviation, pg/g	3,2
Relative standard deviation, %	13
No. of values reported	17
No. of values removed	0
No. of reported non-detects	0




Analyte solution

Congener: PBDE 209

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
3	97	0,037				
4	92	-0,23				
11	97	0,060				
20	82	-0,75				
22	100	0,20				
28	119	1,2				
39	97	0,015				
43	95	-0,079				
62	101	0,22				
76	102	0,28				
79	92	-0,24				
80	98	0,097				
90	91	-0,29				
93	83	-0,69				
100	93	-0,16				
106	96	-0,015				

Consensus statistics

Consensus median, pg/g	96
Median all values pg/g	96
Consensus mean, pg/g	96
Standard deviation, pg/g	8,5
Relative standard deviation, %	9
No. of values reported	16
No. of values removed	0
No. of reported non-detects	0



Analyte solution

Congener: α -HBCD

Lab code	Conc. pg/g fw.	Z-score	Notes	Lab code	Conc. pg/g fw.	Notes
4	499	0,0000				
93	488	-0,11				
106	525	0,26				

Consensus statistics	
Consensus median, pg/g	499
Median all values pg/g	499
Consensus mean, pg/g	504
Standard deviation, pg/g	19
Relative standard deviation, %	4
No. of values reported	3
No. of values removed	0
No. of reported non-detects	0

